



**EU Declaration of Conformity
According to EMC Directive 2014/30/EU**

Applicant's Name : SUMMIT
Applicant's Address : 58, Gaetbeol-ro, Yeonsu-Gu, Incheon, Korea
Manufacturer's Name : SUMMIT
Manufacturer's Address : 58, Gaetbeol-ro, Yeonsu-Gu, Incheon, Korea

Declares that the product:

Product Name : Flue Gas Analyser
Base Model : TPI DC711
Variant Model : -

To which this declaration relates is in conformity with the following standard(s)
or other normative document(s);

EN IEC 61326-1:2021	Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements
EN 55011:2016+A2:2021	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics – Limits and methods of measurement (Exceptions) 30 m Test method clause 7.6.7 Electric welding equipment
EN IEC 61000-3-2:2019	Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)
EN 61000-3-3:2013/A2:2021	Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection

following the provisions of Directive(s);

2014/30/EU Directive 2014/30/EU of the European Parliament and of the Council of 26 February 2014
on the harmonisation of the laws of the Member States relating to electromagnetic
compatibility (recast) Text with EEA relevance.
OJL 206, 6.8.2019, p. 27–30 (BG, ES, CS, DA, DE, ET, EL, EN, FR, HR, IT, LV, LT, HU,
MT, NL, PL, PT, RO, SK, SL, FI, SV)

Korea / 02. 21. 2024

(Place and date of issue) (name and signature or equivalent making of authorized person)

European Contact (for regulatory topics only):